

The Effect of Barium Swallowing on the Reduction of Artifact of the Inferior Wall of the Myocardium

¹Farshid Gheisari, ²Pouya Hekmati, ³Hamid Zare, ⁴Zahra Nasiri,
⁴Azim Kaveh Ahangar and ⁴Mohammad Atefi

¹Nuclear Medicine and Molecular Imaging Research Center,
Namazi Teaching Hospital, Shiraz University of Medical Sciences,

²Shiraz University of Medical Sciences,

³Fars Agricultural and Natural Resources Research and Training Center, AREEO,

⁴School of Paramedical Sciences, Ionizing and Non-Ionizing Radiation Protection
Research Center, Shiraz University of Medical Sciences, Shiraz, Iran

Abstract: Occasionally, gastrointestinal activity in radionuclide scintigraphy could be problematic. Drugs which are labeled by Technetium-99m in myocardial perfusion scanning such as tetrofosmin, sestamibi are released mainly through the hepatobiliary system. Following this release, the digestive system adjacent to the myocardium becomes active which is able to create artifact in the resulting images. The aim of this study is to propose a method that while being safe and inexpensive can provide the least possible changes in the normal myocardial perfusion imaging procedures and also provide better images with higher quality and fewer artifacts. The study sample consisted of 15 cases (6 male and 9 female) aged 40-70 (mean 55) who were referred to our nuclear imaging center and were more likely to develop diaphragmatic artifact on the left ventricle walls of the myocardium, i.e., cases with Body Mass Index (BMI) >25 or those with central obesity. About 30 g of barium sulfate which was solved in 200cc of water at room temperature was administered to the patients in sitting position and immediately they underwent planar scanning of the areas under diaphragm for 5 min. Images of each stage were reviewed and analyzed by the Xeleris 2006 software. All phases of the reconstruction of images, selection of the subdiaphragmatic area, calculation of the count in the selected area and also determination of the radiation activity level in each pixel of the selected area was conducted by this software. According to the results of this study, it can be concluded that there is a possibility of reducing the subdiaphragmatic artifact with this method.

Key words: Myocardial perfusion imaging, barium, artifact, drugs, pixel

INTRODUCTION

About 201 Thallium is recognized as a useful tracer for myocardial perfusion imaging (Fallahi *et al.*, 2014; Pohost *et al.*, 1977). However, due to its physical and biological characteristics, it is not ideal for imaging. Lower photon energy is the major limitation of 201 Tl which causes scattering from overlying tissues. Higher energy of 99 mTc-sestamibi is optimal for gamma camera scintigraphy and it has also less attenuation. Because of the physical advantages of 99 mTc-sestamibi over 201Tl and also more favorable biological characteristics, including rapid lung and liver clearance the development of a myocardial perfusion agent labeled with 99mTc is attractive (Li *et al.*, 1988; Leppo and Meerdink, 1989;

Haghighatafshar and Farhoudi, 2015). Occasionally, some incidental findings and gastrointestinal activity in radionuclide scintigraphy could be problematic (Eftekhari *et al.*, 2010; Haghighatafshar and Khajehrahimi, 2015; Haghighatafshar and Farhoudi, 2015). Drugs which are labeled by Technetium-99m such as tetrofosmin, sestamibi are more favorable radiotracer in Myocardial Perfusion Imaging (MPI) (Ghaedian *et al.*, 2015) and are released mainly through the hepatobiliary system (Higley *et al.*, 1993). Following this release, the digestive system adjacent to the myocardium becomes active which is able to create artifact in the resulting images (Rehm *et al.*, 1996; Karayalcin *et al.*, 1998). High colon activity can prevent the detection of defects in the myocardium inferior wall due to its radiations, especially

in studies with planar images (planar). Accordingly, to reduce the subdiaphragmatic activity or digestive activity and consequently to acquire images with higher quality and accuracy, various studies have been conducted and various methods have been proposed. In these studies, for example, drinking cold or a large amount of water before scanning, eating before scanning, consuming high fat food, drinking juices and administering various drugs have been examined. In this study, we decided to examine the theory that whether by oral administration of barium sulfate prior to scanning, the subdiaphragmatic activity can be reduced. The reason for choosing barium sulfate in this study is firstly, the very good radiation-absorbing property of this material, that on this basis, it is widely used in the other radiological investigations. Another reason is the safety of this material and finally the low cost, ease of preparation and availability of this material at health care centers of our country. In fact, in this study we aim to propose a method that while being safe and inexpensive, can provide the least possible changes in the normal myocardial perfusion imaging procedures and also provide better images with higher quality and fewer artifacts.

MATERIALS AND METHODS

The study sample consisted of 15 cases (6 male and 9 female) aged 40-70 (mean 55) who were referred to our nuclear imaging center. The cases were selected with simple random sampling method among the patients who were more likely to develop diaphragmatic artifact on the left ventricle walls of the myocardium, i.e., cases with Body Mass Index (BMI) > 25 or those with central obesity. Cases who could not swallow barium and cases with contraindication for barium swallowing, i.e., cases who had the possibility or probability of bowel perforation or those who had a history of allergy to barium or cases who were supposed to undergo a CT scan of the abdomen or chest during a few days after the study were excluded from the study. In the studied sample, after obtaining the informed consent from the patients, the intervention took place. The usual method done in this center is the injection of 0.3 mCi kg⁻¹ of the MIBI intravenously, manufactured by Kavoshyar Company of Iran's Atomic Energy Organization and the phase of stress myocardial perfusion scan is carried out by intravenous administration of the Dipyridumole drug by the Gamma Camera (dual head) GE. Xeleris 2006 device on 1-1.5 h after injection. Upon completion of this phase, the research protocol, without the injection of additional doses compared to the conventional method, was conducted a planar (Anterior Projection) picture were made by the

mentioned device for 5 min from the areas under the diaphragm (subdiaphragmatic areas) of the patients. Consequently, 30 g of barium sulfate which was solved in 200 cc of water at room temperature was administered to the patients in sitting position and immediately they underwent planar re-scanning of the areas under diaphragm for 5 min. In the following, according to the routine method performed at the center, with an intravenous injection of 0.3 mCi kg⁻¹ of MIBI material, manufactured by the Kavoshyar Company affiliated with Iran's Atomic Energy Organization, on the day of admission, the patients underwent the rest phase.

Ultimately, the obtained planar images were compared and analyzed before and after using barium sulfate. Results were compared both visually and statistically. The statistical comparison was done. An area of approximately 1400 pixels in the obtained planar images was isolated from the diaphragmatic area under the heart. Then the amount of activity per pixel (Mean Count Per pixel) in the isolated area under the diaphragm was measured and the amounts of activities in each pixel of both images were compared. Wilcoxon test was used for comparison. SPSS 22 software was used for statistical analysis.

RESULTS AND DISCUSSION

The distribution of age, sex and weight of the subjects is presented in Table 1. At the beginning of the study, 80 g of barium sulfate dissolved in 200 cc of water at room temperature but was reduced to 30 g of barium sulfate due to the high concentration of barium and non-cooperation of the patients.

Images of each stage were reviewed and analyzed by the Xeleris 2006 software. All phases of the reconstruction of images, select the subdiaphragmatic area, calculate the count in the selected area and also determine the radiation activity level in each pixel of the selected area was conducted by this software. The results are detailed in Table 2.

- ROI 1 (kc): The amount of the subdiaphragmatic radioactivity in the initial images in kilo kants
- ROI2 (kc): The amount of the subdiaphragmatic radioactivity after taking barium sulfate in kilo kants
- ROI 1 (pix): The selected subdiaphragmatic area in the primary images in pixels
- ROI 2 (pix): The selected subdiaphragmatic area after taking barium sulfate in pixels

To analyze the statistical data, the Wilcoxon signed ranks test was used. The results are provided in Table 3 and 4.

Table 1: The distribution of age, sex and weight of the studied population

| Ages | Sex | Weight (kg) |
|------|-----|-------------|
| 57 | F | 70 |
| 54 | F | 70 |
| 65 | M | 80 |
| 51 | M | 97 |
| 65 | M | 85 |
| 54 | M | 72 |
| 44 | F | 73 |
| 55 | F | 57 |
| 45 | F | 55 |
| 52 | F | 63 |
| 56 | M | 75 |
| 66 | F | 83 |
| 46 | F | 80 |
| 60 | M | 74 |
| 55 | F | 58 |

Table 2: The selected area, the activity level in the selected area and in each pixel of the selected area separately for each patient, before and after using barium sulfate in the planar images

| Roi (kc) | | Roi (pix.) | | Mean count per pixel | |
|----------|--------|------------|------|----------------------|-------|
| 1 | 2 | 1 | 2 | 1 | 2 |
| 125.89 | 84.6 | 1406 | 1406 | 8.95 | 6.01 |
| 153.20 | 118.69 | 1400 | 1404 | 10.94 | 8.45 |
| 183.06 | 130.32 | 1400 | 1404 | 13.07 | 9.28 |
| 121.08 | 88.42 | 1400 | 1400 | 8.65 | 6.31 |
| 113.81 | 94.46 | 1404 | 1406 | 8.11 | 6.72 |
| 148.03 | 121.40 | 1394 | 1400 | 10.62 | 8.67 |
| 193.69 | 151.60 | 1406 | 1406 | 13.77 | 10.78 |
| 161.71 | 111.99 | 1406 | 1406 | 11.50 | 7.96 |
| 161.31 | 120.43 | 1406 | 1406 | 11.47 | 8.56 |
| 181.88 | 123.58 | 1406 | 1406 | 12.93 | 8.97 |
| 339.14 | 265.39 | 1400 | 1404 | 24.22 | 18.90 |
| 176.55 | 134.69 | 1404 | 1406 | 12.57 | 9.58 |
| 236.19 | 193.88 | 1406 | 1406 | 16.80 | 13.79 |
| 210.05 | 163.13 | 1406 | 1406 | 14.94 | 11.60 |
| 270.82 | 196.86 | 1406 | 1406 | 19.26 | 14.00 |

Table 3: The average of activity per pixel and its standard deviation before and after using barium sulfate in the planar images

| Pairl | Mean | N | SD | SE |
|-------|---------|----|---------|---------|
| x_1 | 13.1888 | 15 | 4.28768 | 1.10708 |
| x_2 | 9.9624 | 15 | 3.43826 | 0.88775 |

x_1 : Activity in each pixel in the first images; x_2 : Activity in each pixel after taking barium sulfate

Table 4: The results of Wilcox on test in the study

| Parameters | X2-X1 |
|-----------------------|-----------|
| z | -3.408(a) |
| Asymp. Sig.(2-tailed) | 0.00100 |

Comparing amount of sub cardiac activity in the selected area of images before and after the barium swallow we can conclude that swallowing barium as a material of contrast in the patients' stomachs can significantly reduce the sub cardiac activity in these patients. The result of this reduction in the sub cardiac activity is the quality improvement of the final images (Boz *et al.*, 2001). This reduction can be explained by two mechanisms: The first mechanism is related to the biochemical and physical property of barium, i.e., the high

capacity of radiation absorption that leads to wide consumption of this material in radiological studies, thus we supposed that it can absorb gamma radiations as well as x rays (Iqbal *et al.*, 2004). The second supposed mechanism that can illustrate the sub cardiac activity reduction in this study was the effect of the volume of water administered to each patient. In fact, the patient's filled stomach by 200 cc water can somewhat take the gastrointestinal activity away from the inferior wall of myocardium due to its volume and weight effect. This reduction eventually leads to an improvement in the image contrast and the quality of image interpretation. The reduction of background radiations increases the contrast of the images (Heller *et al.*, 1997). As the whole steps from capturing the first stage images to swallowing barium and taking the second stage images was roughly 15 min, therefore, decreasing the sub diaphragmatic activity due to time passage is negligible. The contrast used in this study was Barium sulfate. The selection of this material was because of its high radiation absorption characteristics. On the other hand, it is known as a drug with very low side effects used in diagnostic procedures. Comparing the first images and the second images visually, significant difference was noted in the second stage images after swallowing the barium. By dividing the total activity by the separated area, the activity was determined in each pixel (Table 2) and the levels of activity on each pixel of the first and second images of the subcardiac area were compared with each other statistically. After performing statistical analysis, the differences in the pictures with $p < 0.001$ are statistically significant as well. According to the results we can conclude that the use of barium sulfate during Myocardial SPECT imaging can significantly reduce the activity of the subdiaphragmatic areas. This reduction in the subdiaphragmatic activity increases the contrast in the images that leads to obtaining higher-quality images of the myocardium. Such as previous study (Haghighatafshar *et al.*, 2015) that indicated the importance of fused Single-Photon Emission Computed Tomography (SPECT)/Computed Tomography (CT) images, SPECT/CT imaging may minimize such artifacts.

CONCLUSION

Using this simple, safe and cheap method during the myocardial nuclear scans, we can improve image quality and reduce the diagnostic errors caused by the background radiations and subdiaphragmatic artifact. It should be noted that further researches and studies in this field in order to determine the most appropriate amount of barium and the best time for barium swallow is needed.

REFERENCES

- Boz, A., A. Yildiz, F. Gungor, B. Karayalcin and M. Erkilic, 2001. The volume effect of the stomach on intestinal activity on same-day exercise-rest Tc-99m tetrofosmin myocardial imaging. *Clin. Nucl. Med.*, 26: 622-625.
- Eftekhari, M., A.F. Esfahani, M. Haghghatafshar, A. Keyvan and D. Beiki *et al.*, 2010. Bone scan with incidental visualization of the entire colon. *Iran. J. Nucl. Med.*, 18: 56-59.
- Fallahi, B., M. Haghghatafshar and F. Farhoudi, 2014. Comparative evaluation of the diagnostic accuracy of 99mTc-sestamibi gated SPECT using five different sets of image acquisitions at stress and rest phases for the diagnosis of coronary artery disease. *Am. J. Nucl. Med. Mol. Imaging*, 4: 10-16.
- Ghaedian, T., S. Mortazavi and M. Haghghatafshar, 2015. Multiple myeloma and abdominal aortic aneurysm on myocardial perfusion raw images. *Clin. Nucl. Med.*, 40: e526-e527.
- Haghghatafshar, M. and F. Farhoudi, 2015. Incidentally visualization of the thymus on whole-body iodine scintigraphy: Report of 2 cases and review of the latest insights. *Med.*, Vol. 94 10.1097/MD.00000000000001015
- Haghghatafshar, M. and F. Khajehrahimi, 2015. Hiatal hernia uptake of iodine-131 mimicking mediastinal metastasis of papillary thyroid carcinoma. *Indian J. Nucl. Med.*, 30: 347-349.
- Haghghatafshar, M., F. Gheisari and T. Ghaedian, 2015. Importance of heparin provocation and SPECT/CT in detecting obscure gastrointestinal bleeding on 99mTc-RBC scintigraphy: A case report. *Med.*, Vol. 94 10.1097/MD.00000000000001325.
- Heller, E.N., P. DeMan, Y.H. Liu and D.P. Dione, 1997. Extracardiac activity complicates quantitative cardiac SPECT imaging using a simultaneous transmission-emission approach. *J. Nucl. Med.* 38: 1882-1890.
- Higley, B., F.W. Smith, T. Smith, H.G. Gemmell and G.P. Das *et al.*, 1993. Technetium-99m-1, 2-bis (bis (2-ethoxyethyl) phosphino) ethane: Human biodistribution, dosimetry and safety of a new myocardial perfusion imaging agent. *J. Nucl. Med.*, 34: 30-38.
- Iqbal, S.M., M.E. Khalil, B.A. Lone, R. Gorski and S. Blum *et al.*, 2004. Simple techniques to reduce owl activity in cardiac SPECT imaging. *Nucl. Med. Commun.*, 25: 355-359.
- Karayalcin, B., A. Boz and F. Gungor, 1998. Early duodenogastric reflux in Tc-99m tetrofosmin myocardial scintigraphy: A case report. *Clin. Nucl. Med.*, 23: 844-845.
- Leppo, J.A. and D.J. Meerdink, 1989. Comparison of the myocardial uptake of a technetium-labeled isonitrile analogue and thallium. *Cir. Res.*, 65: 632-639.
- Li, Q.S., T.L. Frank, D. Franceschi, H.N. Wagner Jr. and L.C. Becker, 1988. Technetium-99m methoxyisobutyl isonitrile (RP30) for quantification of myocardial ischemia and reperfusion in dogs. *J. Nucl. Med.* 29: 1539-1548.
- Pohost, G.M., L.M. Zir, R.H. Moore, K.A. McKusick and T.E. Guiney *et al.*, 1977. Differentiation of transiently ischemic from infarcted myocardium by serial imaging after a single dose of thallium-201. *Circulation*, 55: 294-302.
- Rehm, P.K., F.B. Atkins, H.A. Ziessman, S.E. Green and E.A. Akin *et al.*, 1996. Frequency of extra-cardiac activity and its effect on 99Tcm-MIBI cardiac SPET interpretation. *Nucl. Med. Commun.*, 17: 851-856.